

SECTION 11 53 13 (FORMERLY 11610)

LABORATORY FUME HOODS NOVAGUARD 60 FPM

PART 1 – GENERAL

Summary:

This Specification identifies the minimum material and construction standards that are required to deliver a quality installation of laboratory fume hoods. Fume hoods shall be supplied in accordance with the requirements of this Specification. The fume hoods identified in this Specification shall include the miscellaneous metal panels and other related components as identified on the Drawings and that are necessary for the complete installation.

Hoods shall function as ventilated, enclosed work spaces, designed to capture, confine and exhaust fumes, vapours and particulate matter produced or generated within the enclosure.

1.1 SECTION INCLUDES

- A. Laboratory Fume Hoods

1.2 RELATED SECTIONS

- A. Division 12 Section 36 00, "Countertops"
- B. Division 12 Section 35 53, "Manufactured Metal Casework"
- C. Division 12 Section 32 00, "Manufactured Wood Casework"
- D. Division 22 Section 40 00, "Plumbing Fixtures"
- E. Division 23 Section 30 00, "HVAC Air Distribution"
- F. Division 26 Section 05 00, "Common Work Results for Electrical"
- G. Related Work To Be Performed By Others:
 - 1. Final installation of all plumbing, service and electrical fixtures attached to fume hood or countertop (excluding piping and wiring within fume hoods).
 - 2. Final connection to service lines of all plumbing, service and electrical fixtures attached to laboratory casework or fume hoods. Pressure testing of all plumbing after final connection.

1.3 REFERENCES

- A. SEFA 1-Latest Edition: Laboratory Fume Hoods – Design, Materials, Use and Testing Guidelines

Science Equipment and Furniture Association (SEFA)

- B. ISO 9001:2008 – Quality Management
International Standards Organization (ISO)
- C. ADA (ATBCB ADAAG) Americans with Disabilities Act Accessibility Guidelines Americans with Disabilities Act (ADA)
- D. ANSI/ASHRAE 110, 2016 Methods of Testing Performance of Laboratory Fume Hoods
- E. UL1805 Standard for Laboratory Hoods and Cabinets

1.4 SUBMITTALS

Refer to Section 01 33 00, "Submittal Procedures," for requirements, procedures, etc.

A. Product Data:

1. Drawings shall include data and details for construction of the laboratory fume hoods as well as information regarding the name, quantity, type and construction of materials (such as hardware, gauges, etc), that will be used to complete the project.

B. Shop Drawings:

1. The laboratory casework manufacturer shall furnish shop drawings illustrating the layout and placement of all laboratory casework and fume hoods as well as any products included in this section.
2. Indicate the type and location of all service fittings and associated supply connections.
3. Preparation instructions and recommendations.
4. Storage and handling requirements and recommendations.
5. Installation methods.

C. Selection Samples:

Submit the following:

1. One complete set of metal color chips representing the manufacturer's full range of available colors.

D. Quality Assurance/Control

1. Design Data/Test Reports: Manufacturer shall submit test data and design criteria which are in compliance with the project specifications.
2. Performance: Fume Hoods, Mott Manufacturing "NovaGuard" model, shall be designed to meet or exceed the American Standard for Laboratory Ventilation and the American Industrial Hygiene Association standard as described in ANSI/AIHA Z9.5. This standard of performance

shall be verified through factory testing in accordance with the established protocol as set out by the ANSI/ASHRAE 110 standard.

3. Fume hoods under this section shall be classified to UL standard 1805 which covers, fire, electrical and mechanical hazards. Approval with scope limited to electrical hazards shall not be an acceptable substitute for UL1805.
4. Certificates: All certifications required in the specifications shall be submitted with the original submittal package under separate cover. Certificates must be provided with the signature of a qualified individual of the supplier.
5. Manufacturers' Instructions: Provide manufacturer's instructions for installation and maintenance of all products provided and installed within this section. Instructions shall be available through a QR code adhered to the face of the fume hood.
6. Submit copy of the corrosion resistant label to be attached to the fume hood exterior with condensed information covering recommended locations for apparatus and accessories.

1.5 QUALITY ASSURANCE

A. Manufacturer Qualifications:

1. The following list of information will be provide to the Architect at least ten (10) days prior to the bid opening:
 2. List of manufacturing facilities.
 3. Manufacturer of fume hoods shall have the capability within their facility of performing fume hood tests based on the latest ANSI/ASHRAE Specification 110.
 4. A list of ten (10) installations of comparable stature completed within the past 5 years.
 5. Construction details depicting the materials, sizes and methods of construction.

B. Mock-Ups

1. Area mockups shall be as indicated on the shop drawings. Post bid mockup areas must be priced for disassembly and reassembly and used within the project.
2. Do not proceed with remaining work until installation is approved by Architect.
 - a) Install base cabinet with specified hardware.
 - b) Install fume hood with specified fixtures.

1.6 DELIVERY, STORAGE AND HANDLING

A. Packaging, Shipping, Handling and Unloading

1. Packaging: Products shall have packaging adequate enough to protect finished surfaces from soiling or damage during shipping, delivery and installation.

2. Delivery: Fume hood delivery shall only take place after painting, utility rough-ins and related activities are completed that could otherwise damage, soil or deteriorate fume hoods in installation areas.
3. Handling: Care, such as the use of proper moving equipment, experienced movers, etc., shall be used at all times to avoid damaging the fume hoods. Until installation takes place, any wrapping, insulation or other method of protection applied to products from the factory will be left in place to avoid accidental damage.

B. Acceptance at Site:

1. Fume hoods will not be delivered or installed until the conditions specified under Part 3, Installation section of this document have been met.

C. Storage:

1. Fume hoods shall be stored in the area of installation. If, prior to installation, it is necessary for the fume hoods to be temporarily stored in an area other than the installation area, the environmental conditions shall meet the environmental requirements specified under the Project Site Conditions article of this section.

D. Waste Management and Disposal:

1. The supplier of the laboratory fume hoods are responsible for removing any waste or refuse resulting from the installation of, or work pertaining to laboratory fume hoods; thereby leaving the project site clean and free of debris. Trash container(s) to be provided by others.

1.7 PROJECT SITE CONDITIONS

- A. Building must be enclosed (windows and doors sealed and weather-tight).
- B. An operational HVAC system that maintains temperature and humidity at occupancy levels must be in place.
- C. Adjacent and related work shall be complete.
- D. Ceiling, overhead ductwork and lighting must be installed.
- E. Site must be free of any further construction such as "wet work"
- F. Required casework must be installed accurately and the project must be ready for fume hood installation.

1.8 WARRANTY

- A. Furnish a written warranty that Work performed under this Section shall remain free from defects as to materials and workmanship for a period of two (2) years from date of shipment. Defects in materials and workmanship that may develop within this time are to be replaced without cost or expense to the Owner.

Defects include, but are not limited to:

1. Ruptured, cracked, or stained coating

2. Discoloration or lack of finish integrity
 3. Cracking or peeling of finish
 4. Slippage, shift, or failure of attachment to wall, floor, or ceiling
 5. Weld or structural failure
 6. Warping or unloaded deflection of components
 7. Failure of hardware
- B. The warranty with respect to products of another manufacturer such as, but not limited to, electrical devices and electronics, is limited to the warranty extended by that manufacturer to the hood manufacturer.

PART 2 – PRODUCTS

2.1 MANUFACTURER

A. Acceptable Manufacturer:

- a. Mott Manufacturing Ltd.as distributed by **New England Lab®** (NEL).
- b. Boston Corporate Offices - 1 Arrow Drive Woburn, MA 01801
 - i. (888) 635-2080
- c. Baltimore Office - 2707 N. Rolling Road Suite 110 Baltimore, MD 21244
 - i. (410) 944-7060
- d. North Carolina Office - 2175 Presidential Drive Suite 130 Durham, NC 27703
 - i. (919) 469-8054
- e. For pricing - info@newenglandlab.com
- f. For product questions – Rand Weyler - rand@newenglandlab.com

B. Substitutions:

Must meet all specification requirements and have prior approval.

C. Requests for substitutions:

All requests will be considered in accordance with provisions of Section 01 60 00.

2.2 FUME HOOD MATERIALS

A. Basic Materials

NOTE: A complete list of basic materials is provided here. Not all models use all materials listed.

1. Exterior Panels, Framing Members, and Furring Panels: Cold rolled and leveled mild steel and shall conform to ASTM A366, finished as in Para. 2.3.
2. Screws: Interior fastening devices; stainless steel screws complete with corrosion resistant plastic caps.

3. Upper front Panel: Minimum 20 Ga (0.9mm) mild steel, finished same as exterior panels.
4. Lower Foil / Flush sill: Type 316-4 stainless steel powder coated.
5. Safety Glass: Laminated type 6mm (1/4") thick as per Section 11 53 00.
6. Sash guides: Track shall be corrosion resistant polyvinyl chloride (PVC).
7. Sash Chain: #35 hardened carbon steel roller chain.
8. Sprocket system for Sash Chain: Hardened sprockets with one full width shaft per sash running in ball bearings.
9. Sash Pull: Type 316, 18 Ga (1.2mm) thick stainless steel with an AISI #4 satin finish.
10. Baffle support brackets: Fiberglass reinforced polyester thermoset resin of 3/16" (5mm) thickness.
11. Duct Stubs: Type 316, 18 Ga (1.2mm) stainless steel.
12. Light Switches: Light switches shall be black in color, commercial spec grade or higher and shall be UL and CSA approved.
13. Electrical receptacles: Electrical receptacles shall be black in color, commercial spec grade or higher and shall be UL and CSA approved.
14. Cover Plates: Electrical cover plates shall be black, nylon and UL and CSA approved.
15. Lighting shall be provided by LED module and driver located on the fume hood roof. Provide a 6mm (0.236") safety glass panel on hood "roof", sealed to isolate the lighting fixture from fume chamber. The LED module and driver combination shall be rated at 5900 lumens with a color temperature of 4000K. One nominal 22" (559mm) module shall be provided for each hood up to 6 foot in size with 8 foot models having two. Average interior illumination levels within the fume chamber shall be 80 foot candles minimum. Efficiency of light shall be 120 lumens per watt, and life expectancy of 50,000 hours.

B. Fume Hood Liner

1. FRP:
 - a) Hood linings and baffles shall be fiberglass reinforced polyester thermoset resin of 3/16" (5mm) thickness.
 - b) The fiberglass reinforced polyester panel shall have a minimum flexural strength of 15,000 psi (103,400 kPa), with a flame spread of 25 or less as per ASTM #E84.
 - c) Final appearance shall be smooth and white in color.
2. PVC type:

- a) Hood linings and baffles shall be Poly Vinyl Chloride of 1/4" (6mm) thickness. Final appearance shall be smooth and white in colour.
 - b) Hood linings and baffles shall be Poly Vinyl Chloride of 1/4" (6mm) thickness with square corners. Final appearance shall be smooth and white in colour.
3. Stainless steel:
- a) 316 s/s square corners.

C. Fume Hood Furring Panels

1. Where called for, provide matching furring panels to enclose the space between top edge of fume hoods and the finished ceiling.
2. Panels shall be flanged, notched and reinforced where required to form a well-fitted enclosure, free from oil canning. Secure panels using cadmium-plated, self-tapping screws; panels shall be removable for maintenance purposes.
3. Front of furring panel shall match fume hood to which it is connected.

2.3 BENCH FUME HOOD CONSTRUCTION

- A.** Fume hood superstructure shall of the restricted bypass, high performance type and be double wall construction consisting of an outer shell of sheet steel and an inner hood liner. Double wall shall house and conceal steel framing members, attaching brackets and remote operating service fixture mechanisms. Overall double wall thickness; 4-3/4" (121mm) maximum. Hood exterior shall be 54.25" in height and 36" in depth.
- B.** Front double-wall posts including sash track shall slop rearward at an angle of 5 degrees. Posts shall accept up to five plumbing fittings per side, one electrical duplex per side, light switch and optional monitor alarm where indicated on drawings. Posts are only punched for plumbing fittings as ordered. Electrical outlets and light switch shall be factory-wired and terminate at a junction box on roof of hood. All electrical components shall be UL listed/classified.
- C.** Fume hood side walls shall be nominally 4.75" (121mm) thick and shall provide space for routing of plumbing supply lines and electrical cables. In addition, each side wall shall contain a 3.5" (89mm) nominal deep recessed alcove to maximise available workspace. The vertical and top sides of the alcove shall be available for installation of plumbing service outlets to prevent intrusion of same into main working chamber. If specified on project drawings, the alcove may contain an optional laminated safety glass viewing window on one or both sides.
- D.** Exterior panel members shall be fastened by means of concealed devices. Exposed screws are not acceptable.
- E.** Provide access to remote-controlled fixture valves concealed between walls through removable panels on hood exterior and access panels on both inside liner walls. Assemble hood superstructure, fasten and connect inner and outer frame into a rigid self supporting entity.

- F. Front Panel shall be contoured powder coated steel and shall cover the roof area of the fume hood and extend nominally 3" (75mm) below the fume hood roof line covering the top edge of the upper vision panel glass.
- G. Vision panel glass shall be tinted or clear (as specified on project drawings) laminated safety glass and shall cover the upper portion of the fume hood front opening between the top edge of the sash glass and the roofline of the fume hood.
- H. Fume hood sash shall be full view type providing a clear and unobstructed side to side view of fume hood interior. In combination with the stationary vision panel glass, vertical viewing area shall be a minimum of 44" (1397mm) above the work surface. Sash shall be laminated safety glass. Bottom and side sash rails shall be 18 Ga (1.2mm) powder coated stainless steel. Bottom rail shall be an integral, formed, full width, flush pull and shall be anchored on each side to sash chains at bottom. A chain, bearing and shaft counter balance system shall be used for vertical operation of sash and prevent jamming to permit one finger operation at any point along full width sash pull and to maintain sash at any position below working height without creep. Sash system shall be designed to prevent sash drop in the event of chain failure. Superstructure shall have a single sash and counter balance system. Sash shall open and close against rubber bumper stops. Maximum sash opening for setup is 28.5" (724mm).
- I. Hood sash shall rise into an enclosure box to assure a leak free chamber.
- J. Self-closing sash: A mechanism shall be provided which automatically lowers the sash to the chosen working height (height shall be 18" (457mm) unless specified at time of order). A latch shall be provided to hold the sash fully open for setup/teardown of experiments. Below the chosen working height, the sash shall be neutrally balanced and function as a conventional sash.
- K. Hood shall be restricted bypass, reduced face velocity design.
- L. Perimeter of sash opening shall have a lower flush sill containment trough and an angled side and top with angled edge toward hood interior. Air shall enter under the bottom flush sill through a nominal 1" (25mm) by-pass when the sash is in the closed position. Bottom foil shall be removable without the use of special tools. Sash shall close on top of flush sill.
- M. One-piece baffle shall provide controlled air vectors into and through the fume hood and be fabricated of the same material as the liner. Provide exhaust slot on the bottom edge of baffle.
- N. Design fume hoods to minimize static pressure loss with adequate slot area around the baffle. Measured average static pressure loss reading taken three diameters above the hood outlet from four points, 90° apart, shall not exceed the following values based on 60" (1524mm) wide hood:

Face Velocity		Measured Static Pressure Loss	
60 F.P.M.	(0.3 m/s)	0.10"	(24.9 Pa)

- O. Airflow Requirements: NovaGuard fume hood is designed to function with the following exhaust volumes when operating at 60 feet per minute face velocity with a vertical sash opening of 18" (457mm)

Hood Size	Exhaust Volume Requirements (cubic Feet Per minute)
4 foot (1220mm)	289 cfm
5 foot (1524mm)	379 cfm
6 foot (1829mm)	469 cfm
8 foot (2439mm)	649 cfm

- P. Electrical convenience duplex outlets shown mounted on the face of fume hoods shall be installed in front posts and pre-wired to a junction box mounted on top of fume hood superstructure. Electrical devices shall be UL classified/listed. Type and quantity shall be as shown on project drawings.
- Q. Attach corrosion resistant labels to units as specified in Para. 1.4.D.4

2.4 FUME HOOD EXTERIOR FINISH

- A. Coating Performance data is available in Appendix 1.

2.5 AIR FLOW MONITOR / ALARM

- A. TEL AFA 1001 Mk3 digital airflow alarm or equivalent shall be provided where shown on project drawings.

PART 3 – EXECUTION

3.1 INSTALLATION

- A. In addition to requirements of Section 11 53 13, install fume hoods in positions shown, align and set level with levelling devices.
- B. Work in close cooperation with allied trades installing ductwork, wiring and other services.
- C. Apply small bead of sealant to junction of fume hood counter top and adjacent hood liner.
- D. Turn over to Mechanical Trades, service fitting remote control rods and valves for installation to fume hood superstructure and service lines.

END OF SECTION